

## Calibration Certificate / *Kalibrierschein*

issued by the calibration laboratory / *erstellt durch das Kalibrierlaboratorium*

**Deutsche WindGuard  
Wind Tunnel Services GmbH**



Deutsche  
Akkreditierungsstelle  
D-K-15140-01-00

Member of / *Mitglied im*  
**Deutschen Kalibrierdienst**



Calibration certificate  
*Kalibrierschein*

Calibration mark  
*Kalibrierzeichen*

2121063
D-K-
15140-01-00
04/2021

<b>Object</b> <i>Gegenstand</i>	3D Sonic Anemometer
<b>Manufacturer</b> <i>Hersteller</i>	Gill Instruments UK-Hampshire S041 9EG
<b>Type</b> <i>Typ</i>	1590-PK-020/W
<b>Serial number</b> <i>Fabrikat/Serien-Nr.</i>	211101
<b>Customer</b> <i>Auftraggeber</i>	Gill Instruments UK-Hampshire S041 9EG
<b>Order No.</b> <i>Auftragsnummer</i>	M52803
<b>Project No.</b> <i>Projektnummer</i>	VT210334
<b>Number of pages</b> <i>Anzahl der Seiten</i>	5
<b>Date of Calibration</b> <i>Datum der Kalibrierung</i>	06.04.2021

This calibration certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI).

The DAkkS is signatory to the multilateral agreements of the European co-operation for Accreditation (EA) and of the International Laboratory Accreditation Cooperation (ILAC) for the mutual recognition of calibration certificates. The presented results relate only to the calibrated object. The user is obliged to have the object recalibrated at appropriate intervals.

*Dieser Kalibrierschein dokumentiert die Rückführung auf nationale Normale zur Darstellung der Einheiten in Übereinstimmung mit dem Internationalen Einheitensystem (SI).*

*Die DAkkS ist Unterzeichner der multilateralen Übereinkommen der European co-operation for Accreditation (EA) und der International Laboratory Accreditation Cooperation (ILAC) zur gegenseitigen Anerkennung der Kalibrierscheine. Die dargestellten Ergebnisse beziehen sich nur auf den kalibrierten Gegenstand. Für die Einhaltung einer angemessenen Frist zur Wiederholung der Kalibrierung ist der Benutzer verantwortlich.*

This calibration certificate may not be reproduced other than in full except with the permission of the issuing laboratory. Calibration certificates without signature are not valid. This calibration certificate has been generated electronically.

*Dieser Kalibrierschein darf nur vollständig und unverändert weiterverbreitet werden. Auszüge oder Änderungen bedürfen der Genehmigung des ausstellenden Kalibrierlaboratoriums. Kalibrierscheine ohne Unterschrift haben keine Gültigkeit. Dieser Kalibrierschein wurde elektronisch erzeugt.*

Date  
*Datum*  
  
06.04.2021

Head of the calibration laboratory  
*Leiter des Kalibrierlaboratoriums*

Dipl. Phys. Dieter Westermann

Person in charge  
*Bearbeiter*

Kai Schuster, B. Eng.

<b>Calibration object</b> <i>Kalibriergegenstand</i>	3D Sonic Anemometer										
<b>Calibration procedure</b> <i>Kalibrierverfahren</i>	IEC 61400-12-1:2017										
<b>Place of calibration</b> <i>Ort der Kalibrierung</i>	Wind tunnel 2 of Deutsche WindGuard Wind Tunnel Services GmbH, Varel										
<b>Test conditions</b> <i>Messbedingungen</i>	<table><tr><td>wind tunnel area</td><td>10000 cm<sup>2</sup></td></tr><tr><td>anemometer frontal area</td><td>220 cm<sup>2</sup></td></tr><tr><td>diameter of mounting pipe</td><td>0.0 mm no mounting</td></tr><tr><td>blockage ratio <sup>1)</sup></td><td>0.022 [-]</td></tr><tr><td>software version</td><td>9.1.0</td></tr></table> <p><sup>1)</sup> Due to the special construction of the test section no blockage correction is necessary.</p>	wind tunnel area	10000 cm <sup>2</sup>	anemometer frontal area	220 cm <sup>2</sup>	diameter of mounting pipe	0.0 mm no mounting	blockage ratio <sup>1)</sup>	0.022 [-]	software version	9.1.0
wind tunnel area	10000 cm <sup>2</sup>										
anemometer frontal area	220 cm <sup>2</sup>										
diameter of mounting pipe	0.0 mm no mounting										
blockage ratio <sup>1)</sup>	0.022 [-]										
software version	9.1.0										
<b>Ambient conditions</b> <i>Umgebungsbedingungen</i>	<table><tr><td>air temperature</td><td>(20.7 ± 0.4) °C – (21.0 ± 0.4) °C</td></tr><tr><td>air pressure</td><td>(1005.9 ± 0.4) hPa – (1006.2 ± 0.4) hPa</td></tr><tr><td>relative air humidity</td><td>(29.7 ± 5.0) % – (30.0 ± 5.0) %</td></tr></table>	air temperature	(20.7 ± 0.4) °C – (21.0 ± 0.4) °C	air pressure	(1005.9 ± 0.4) hPa – (1006.2 ± 0.4) hPa	relative air humidity	(29.7 ± 5.0) % – (30.0 ± 5.0) %				
air temperature	(20.7 ± 0.4) °C – (21.0 ± 0.4) °C										
air pressure	(1005.9 ± 0.4) hPa – (1006.2 ± 0.4) hPa										
relative air humidity	(29.7 ± 5.0) % – (30.0 ± 5.0) %										
<b>Measurement uncertainty</b> <i>Messunsicherheit</i>	<p>The expanded uncertainty assigned to the measurement results is obtained by multiplying the standard uncertainty by the coverage factor <math>k=2</math>. It has been determined in accordance with EA-4/02 M: 2013. The value of the measurand lies within the assigned range of values with a probability of 95%.</p> <p>The reference flow speed measurement is traceable to the German NMI (Physikalisch-Technische Bundesanstalt) standard for flow speed. It is realized by using a PTB owned and calibrated Laser Doppler Anemometer (Standard Uncertainty 0.2 %, <math>k=2</math>)</p>										
<b>Additional remarks</b> <i>Zusätzliche Anmerkungen</i>	Orientation: 180° North alignment via transducer support arm										
<b>Revision</b> <i>Revision</i>	0										

**Calibration result**  
*Kalibrierergebnis*


Reference	Combined	Test item	Test item	Test item
Air velocity	Unc	Speed	Direction	Vertical
m/s	m/s	m/s	deg	m/s
3.903	0.040	3.847	180.000	-0.032
5.909	0.040	5.824	180.000	-0.051
7.971	0.040	7.865	180.000	-0.071
9.983	0.040	9.852	180.000	-0.080
12.011	0.048	11.848	180.000	-0.081
13.944	0.056	13.741	180.000	-0.085
15.979	0.064	15.741	180.000	-0.087
15.048	0.060	14.823	180.000	-0.085
13.006	0.052	12.813	180.000	-0.086
10.987	0.044	10.831	180.000	-0.082
8.952	0.040	8.823	180.000	-0.071
6.950	0.040	6.844	180.000	-0.059
4.910	0.040	4.821	180.000	-0.047

<b>Statistical analysis</b>	Slope	1.01463 (m/s)/(m/s) ±0.00067 (m/s)/(m/s)
	Offset	0.0010 m/s ±0.007 m/s
	Standard error (Y)	0.010 m/s
	Correlation coefficient	0.999998

**Remarks**                      The calibrated sensor complies with the demanded linearity of MEASNET



**Graphical representation of the result**  
*Grafische Darstellung des Ergebnisses*



**Photo of the measurement setup**  
*Foto des Messaufbaus*



**Sensor config during calibration**

*Sensorkonfiguration während der Kalibrierung*

Serial Number : W211101  
---- SW VERSION -----  
Anemometer Firmware Version: 2329-701  
---- HW DETAILS OF ANEMOMETER -----  
Detected Anemometer Class : WindMaster 20Mhz  
Number of AXIS on this unit : 3  
DAC resolution of attached unit : 12  
Number of DAC channels : 0  
Maximum sampling rate : (8) P8 20 Hz  
---- USER CONFIGURATION OF ANEMOMETER -----  
Baud Rate : 19200  
Output Message Format : M2 Polar  
Output Message Velocity Units : m/s  
Output Message Terminator : <cr><lf>  
Retries : ON  
Instantaneous Sampling : OFF  
Factory Calibration : ON  
Physical Communications : Auto  
Resolution settings : Normal resolution  
Alignment settings : North to spar  
Averaging settings : Averaging inactive  
Output rate : P1 1 Hz  
Polar angle minimum magnitude: 0.050 m/s  
Power-On Message : Display Power On Message  
ASCII Format : Comma Separated (CSV)  
SOS and Sonic Temperature display : Neither  
Analogue output scaling : 5m/s  
Unit Identifier : 'Q'  
Synchronised Polling : Off

- End of document / Ende des Dokuments -



Test Report issued under the responsibility of:

DEUTSCHE  
WINDGUARD

TEST REPORT  
IEC 61400-12-1 Annex F  
Anemometer Calibration Certificate

IECRE Report Number ..... : IECRE.WE.TR.AC.21-19007-R0

RETL Calibration Certificate ..... : 2121063

Date of issue ..... : 06.04.2021

RE Testing Laboratory ..... : Deutsche WindGuard Wind Tunnel Services GmbH  
(Name & Address) D-26316 Varel

Applicant ..... : Gill Instruments  
(Name & Address) UK-Hampshire S041 9EG

Test item description ..... : 3D Sonic Anemometer

Manufacturer ..... : Gill Instruments

Model/Type reference ..... : 1590-PK-020/W

Ratings / Serial number ..... : 211101

Tested by (name, function, signature) Printed name/function Signature  
..... : Kai Schuster, B. Eng.  
Operator 

Approved by (name, function, signature) Printed name/function Signature  
..... : Dipl. Phys. Dieter Westermann  
Head of calibration laboratory 

Copyright © 2018 IEC System for Certification to Standards relating to Equipment for use in Renewable Energy applications (IECRE System). All rights reserved.

This publication may be reproduced in whole or in part for non-commercial purposes as long as the IECRE is acknowledged as copyright owner and source of the material. IECRE takes no responsibility for and will not assume liability for damages resulting from the reader's interpretation of the reproduced material due to its placement and context.

If this Test Report Form is used by non-IECRE members, the IECRE logo and the IECRE report number shall be removed.

This report is not valid as a Test Report unless signed by an approved RE Testing Laboratory.

General disclaimer:

The test results presented in this report relate only to the object tested.

This report shall not be reproduced, except in full, without the written approval of the Issuing RE Testing Laboratory (RETL). The authenticity of this Test Report and its contents can be verified by contacting the RETL, responsible for this Test Report.